

CONFIDENTIAL

01-Jun-16

Barts and The London  
NHS Trust



## PEEZY MIDSTREAM QUALITY IMPROVEMENT AUDIT

Interim results

	Historical		Peezy	200
	Quantity	Percentage	Quantity	Percentage
Flow Cytometry indicates no culture needed	37,265	38.38%	51	77.27%
Heavy mixed growth	6,935	7.14%	0	0.00%
Profuse mixed growth	3,643	3.75%	1	1.52%
Moderate mixed growth	6,279	6.47%	0	0.00%
No significant growth	13,577	13.98%	3	4.55%
No growth	7,343	7.56%	10	15.15%
Urine cultured and yielded a significant result	22,044	22.71%	1	1.52%
<b>Average contaminated sample rate (heavy, profuse and moderate mixed growth)</b>		<b>17.36%</b>		<b>1.52%</b>
<b>INTERIM TOTAL</b>	<b>97,086</b>		<b>66</b>	